6/29/09 Breath Alc. Pgm

- ✓ Count calendar - initial changes
- ✓ range on DMT Sim .095 out of range - range = .095 - .105
- ✓ acetone cal. sol'n check
- breath flow conversion issue
- blank error problem
  proximity of exhaust
- detectors & other parts - are here
- Tower/ Host software present but not working - Bertrand Sands
  Empire Tool

6/29/09 Breath Alcohol

- checking RPCs for the "drift" instruments
- Shitt Co. PMs - coordinate w/ communication patch update
- Sim Solns needed .01 field
  .01 calibration
  .02 acetone in .10 EtOH