2/20/09
Alcohol Meeting

- BAC Status
  - Northfield 2/17 + OCSD 3+ weeks down
  - APM in SE State
- DMT APM Franklin/GI in next few weeks. Connect Phones
  Will follow APM SOP as written and possibly modify as needed after
  - Contact Holly @ SACPD re: RFI Problems Emailed 2/20/09
  - SACPD RPC note "VDHL noticed detector voltage continues to climb. VDHL is monitoring situation." Done 2/20/09
- DMT Software
  Waiting on patch from NPAS
  VDHL (Bob) must send email re acceptance range. Call Scott for Comm

DMT / Dräger Testing - waiting on filters for 10/7/08 to finish 1st round of interferent testing. Will continue testing

Dräger - Vol will need tank, RFI immune, Test per air protocol, Acetone no effect

Meet monday IR lab + Survey

3/9 remote update / station (1-5?) / summary reports

- Scott to send software 3/10
- 101708 Def filters again?
- Survey comments
  - email chiefs / Sups
  - post link on Spilman
  - poster

6 month exp on TMX

BAC

SE corner of state done
Manchester in-house
- Bob will get clarification on what to do w/ BAC stored data if we find it
- Steve heard from Mossier that some of the instruments (Casella or VSI-Rent) might be collecting data but unknown which unit, no TSI
- BAC decommissioning? Not yet, RSD/STI will make list
- Replace VCJTC units? Not yet
- OCSD got new unit
-

101708 - Install new block + retest STI will have ready for Wed.

Check Sup list

STI to do Frank APM's test of 3/16

3/16
ALB will look @ DMT software tomorrow
1017 interference test. If failed RTV
if passing finish testing
finish testing on other 2 DMT & Dragger
by 3/31

Dragger email - cal factor response
To probe, Software upgrade

Discovery Responses -

→ P-Alc-101 Sample Prep ← ALB to find &
finish TMX, IS

12 BAC's to retire SH to
make list w/ justifications
SH to replace all VPA units by June

Case review 3/27ish

ALB to forward tech info to SH
Check interference doc.
SH to call NPAS re: Schematics 2006?

NPAS docs to have dates added for
receipt. No watermarks, Stamp Do Not Copy,
RJD to call NPAS re: disseminating manuals

3/23

ALB to finish shopping list

SH to cal/cert 101708

DMR wrote up Draeger justification, recommends unit for next purchase

P-Alc-203 Sim Solution needs revis.
RPC-for BAC MIA again?

AB
DR
- New instrument fantasy features
- Software protocol

4/6

- DM working group available as webinar from ~~Anchorage~~ IACT (5/09)

- SH to send time change reminder email.

- SH / RJD to schedule decommissioning of BAC units

DMR / ALB to finish DMT / Draeger testing this week

123106
~~123106~~ blown modem

ALB - clarify "clear memory"
clarify download dup data

Call SC re: Comm capabilities
file structure etc.

Brett Baker 803-896-7583
Debbie Banks

Brett ~~VSP~~ - 1/17 weird ticket no breath results
St. J 3/14 .118 1st .000 2nd Stuck 5-way

Bruce Smith

SC 153

All in field for ~1 month

- filter wheel issues - rubber wheel had flat spot
- Working on communication software to Poll units, should have by end of week - Polling tomorrow
- ~20% didn't time change

1st week of Jan rolled out on site

pt. Cal, A+P, diag, non-DUI test,
.08

RFI 3point Cal in lab .02 .15 .4

- ✓ cell phones

Ask about boards losing RFI

Bad glue on Solenoid

Commerror replacing cable on pcb

Pump error lengthen trigger time

APM - every 9 weeks (12 week deadline)

4K officers 30 units 90 officers/day

If Ø alc no interference

SC has comm downloads

love Scott ?!?!

Had to pay extra to parse data

DMT's done testing (lin/Int)
101708 + draeger need vol
105108 + draeger test ??

~~123~~ 123106 modem?

- ? Chitt units telephone?
- Scott finishing comm
- Should we poll?

Blood discovery response check list
~~Chris Gilmore~~
217-245-2183

Call NPAS
tank / password.

- John never ordered tank, Dave will
- ILMO shipping gas
Placed 3/24

Prep DMT's for VPA training

Meet w/ Bob re: Discovery document

09-19-100 high?
09-18-100 low

1 linearity
methanol
Acetone
✓ out dry gas

6/2

Drop off in curve 4-5 Sr-AC

BPB w/o 6/8 + 6/15

Call Scott

- Drager pack-up
- uniform data packet
- sim bottles

due 6/26
- Cal w/ current Ace + test w/ int. 3-5 tests on >3 inst
- Cal w/ new Ace + test w/ int. 3-5 tests on >3 inst.

2 replacement + w/of 6/8
2 w/o 6/15

Essex 122406
FCSO St Alb recal

Depletion Study - run till gone
- long term study

6/15

logbook / Cal + Cert training
DMT Software updates
detectors

laptop pw DMT2009
- STI to find DM Host Disk
- tower requiring PW

Finish software testing of DMT.exe
test DMTech
test Parsing software

Burn-in - STI to find out how burn in is done, what graphs mean. Lab alc will determine if in-house burn in is still necessary.

8 detectors, RMA on def detectors etc
PCB if needed, Breath tube

6/29

Initial Changes on court calendar

Volume Conversion
- get "real" ticket + excel ticket

Answer to ≥ blank error
→ find "Blank error" NPAS issue

software error list to Scott 6/30

- Install new detectors in bench units,
  ✓ June RPC to determine field units

✓ DMT log out ✓

will to call ~~Scott~~ Chris + get pic / dwcmp of exhaust

SH to order 80 washers to fix sticky 5-ways

Tie update versions of fix lists

Fri 9:30 Mtg re Status Codes

7/27 SAC. SUP Manual

Still only 1 BAC!

- ✓ Int solns for new DMTs
- Review Power-up Δ

- 15 new received & accepted 9/30

✓ Call Eliza Lewis for printers

- ~~✓ data range on summary~~
Add test sheet for weekly printer

~~- heated sim tubes~~

✓ "Suck back Error" does not fail RPC

called Carolyn
- Update Screen shots
- Call Cindy for # Wash | Add. FLO, Constables, aux
- training site in Wash

$ - training materials
books
solution
projector ? ~ $600 - 1000 retail

1/29/10

2/1 + 2/2 DMT Supt training @ VPA (SH & Alex)
2/1 DMT SA Training (ESD & DMR)

Possible need for DMT op training @ VPA
for DMV/Wind/Wash

Order for final 40 pending
1st 20 here (?) Mid Feb?
most likely March
- Training still on for April

Order submitted for refurb. pending
(SH)
DMR to inventory sim solution

Windham deploy 2/5, 2/8 & 2/9

* 2/5/10
Next 40 order went to purchasing
Cindy confirmed Hartford / Derby
waiting on Lamoille, Addison, Bennington

* Field readiness check-list for
new units (SH to make)

* Put tags on bad sims ✓

DMR + ALB to finish field cal/cat/ready of 6

2/5
SH to fix units Monday, Wind 3 on 2/9
ALB trial on 2/9. Dover/Wilmington 2/10
DMR to Bratt 2/11

, to call
es NPAS
ant?
oards?

Decomp 88. Keep 6 after deployment

Trash pick-up w/of 2/22 ✓ email Cara

SH to ask NPAS who else uses BAC's & do they want parts

DMR updated white board + electronic but hasn't gotten to checking TSI's

DMR to look into Access database for DMT records

Summary sheet for units to track locations

DM log + documentation training to be hosted by ALB include Cara for monthly updates

2/16
Call w/ Scott
- data stored as PPM then divided by 2605 ~~for~~ so result has problem in 32nd decimal place. If round in 6th decimal, then truncate to 3 it will fix problem

In writing

hd 2/15

SH to call Bratt re: time pressure

Bratt PD DMT changed date/time when turned on, DMR adjusted — installed w/ 2/14 .103, called .101, OOR on 2/14 @ .096.

SH got Winooski PD ??

Fabrication

☆ → Call IT re: access to DM Host PC ✓

☆ → Why defaults date/time back when off/on?

Wilmington PD, Darcy going to Bratt this week for trial, will switch 5-way

Suck-back / one-way valve ? Should we change it?

Update software SOP

Find IT sign-off for thumb drive

Software rev VT1.6

2/16

ALB checked remote connection to Williston/Winooski. Williston works. Stt lied about Winooski connection impossible. Stt claimed that he copied + pasted the phone # from DMT log into the dialing software + accessed Winooski's DMT and changed the date/time. Copy/Paste does not work w/ dialing software, and the phone # we have for Winooski includes a * which DMTalk does not accept.

6520390

~~1/1/2006 - default fall back date~~
1/1/2008

2/17/10 Washington County SO had 5 failing RPC's ~~with~~ prior to one passing. There are no notes on the rpc's regarding the failures or what may have caused them. SH did not follow-up w/ WCSO re: failures. RJD instructed that follow-up was not necessary even though VDHL has no idea why the unit failed if it was an instrument problem or operator error.

2/22

Barre - screen not working due to being off-centered, NPAS looking into rubber gromit (?) to prevent

Wilmington - 5-way "fixed"

104709 - repaired

Richmond - recal defective detector

Stt will get sim ready for '06 unit for Wed Sup class (3 DMT's for class)

Stt claims 10 sims for DMT

DMR to inventory sim soln b/c Stt is incompetent

Check w/ Scott re: rounding explanation

2/26

1047- SH testing 5-way valve
testing depletion 100 runs on
Sim w/ ∅ depletion
1038- .095 out of .100 on install
.092 RPC during

Interference testing w/ Ashley

Richmond- voltage unstable, detector
loose. Detector drifting
again lamp loose, voltage
remains unstable 4mV over
~10sec. Replaced detector,
new detector not able to
optimize
- will get loaner (1038?) if
not fixed by 3/8

AeB to cover court 3/3 + 3/4

3/8/10
· Sup 3/15 1pm
· Cal / Cert log books

104809 Mtplr. remove coozie) SH
fax summary rpt.
run check-in first

Call Wash to SA + get summary reports

~~Check off~~ log re: Carolina Alvarado

Get Ed for blood run

- Review Procedure for DMT reports

- MT EtOH Samples
- Sim Solution
- Schedules (white board)
- CAP PT's
- DMT files
- Testing update
- Training update
- filing update
- Sim Soln.

- Bobs shelf SOP - P-ALC-001
- Middlesex

Darcy testimony 1/12/11

- emails from her regarding concerns are "missing"

- interference testing 80% acceptable

- Wash county interference testing concerns

- "No case specific IR's"

- too many interference errors now

- met interference in testing then failed @ field cal

- T° issues

-

- Connor Stewart 4770-12-10
432-12-10 CNCS

4/25/11

- DMT lab will be locked with combo

- 100170 - SM Cal/cert + test close out TSI passing Meott test

VSP - Royalton
- 100151 - Tolerance check off
  Corrective action
  manual inspection of tickets to flag errors
  TSI
  SA had warning in 10/10

  Check all DMTs for tolerance/options
  SH will draft email to have all agencies print options & send to us
  ALB will meet w/ Ed for CA

- Rutland PD TSI - still troubleshooting working w/ NDAS

- Hardwick PD TSI - Pending

SH going to Windham county for APMs 4/26/11 If NEP detector unit will be pulled

ACB will add tab in DMT log to document ~~to docum~~ detector status and ~~PCB~~ Controller Board status on APM tab

- SH to go to Brighton to fix ~~Sim~~ reset spelling + Essex So to fix Sim + do APM, pull if NEP

SH will get APM Schedule - Table

RPC instructions
Spl Acc test > table

7/11/11

- 3im Soln log sheet
- WIP / Review folders need update
- Software
- SH working on SOP manual
- AB working on QP manual
- files - make own folders
  lab_a1c for incoming
  - Inter office
  - inter agency
  - NPAS
  - legal
  - Blood

  Monthly

  DMT_emails June 2011
- SOP's - keep moving on review

7/18/11

- ~~Software~~ RPC loosing flag
  GI, LCSO? St. Alb, + others

SH says this may be happening more often

- ~~Software~~
- SH needs to do June .PST today
- Lab- a/c for June .PST

- Software
  - New S1 coming for 100152
  - Ask Scott about blanking - we aren't getting blank errors anymore
  - Call/ent off/on flagging still not working? test
  - Finder/USB not working well
  - SH to test Mon. AB on Tue/Thu SM on Wed. SH/KK Fri

SOP/Forms - Get them going

- DPS put out a DUI program assessment bid for a reviewer. Breath, blood, legal

- Kirk to clean up pending TSIs

8/8/11

5 units in house after NPAS
Fresh Start will get power today

Software

2 people in room at all times. Must
be locked

Scanner coming on Thrs IT will
teach us how it works

Sim soln ready

SH to pull 1 St Jay + 1/2 of Windham
city