DataMaster Monthly Update Check-list

July 2015

Department/Agency: VSP Williston DM serial #: 100132

Date: 8 / 7 / 15

	Provided	None this month
*Operators Logbook *(Monthly)*	✓	
*Maintenance Logbook *(Monthly)*	✓ No maint. Perfor med	
*Breath Summary Report *(Monthly)*	✓	
*Status Summary Report *(Monthly)*	✓	
Calibration *(Monthly as needed*		✓
Certification *(Monthly as needed)*		✓
Installation *(Monthly as needed*		✓
Annual Preventative Maintenance		✓
Routine Performance Check *(3Xs year)*		✓
Simulator Solution Change *(Monthly)*		✓
Other Reports		

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

VSP Williston
Department/Agency

100132
DataMaster DMT Serial Number

	Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2	
APRIL 2015	04/16/15	Barone	Tomlinson	.095		
	04/18/15	NADEAU	BRYAN G. COOKE	.098		
	04/25/15	LEISE	ALEXANDER MCGOWAN	.098 ① .098 ②		
MAY 2015	05/02/15	NADEAU	ROBERT SMITH	.094		
	05/05/15	NADEAU	CHAD E. LANG	.095		
	05/09/15	LEISE	MICHAEL M. BURNHAM	.093	.093	
	5/11/15	Metayer	Daniel King	.096	.095	
	5/13/15	LeBlanc, M	DOMINA, PAUL L.	DRUG		
	5/17/15	LeBlanc, M	Taylor, Michael	.098		
	05/29/15	Barone	Hackborth, Neil	.094		
June 2015	06/01/15	LEISE	SAWER, MICHAEL C.	.098		
	6/4/15	R. Driscoll	TEST / ARM	.095	—	
	6/26/15	J. Fox	TERRY SANTEE	.098		
	6/30/15	B. Connor	Adam Roy	.096	.098	0.098
July 2015	7/3/15	Metayer	William Gunther	.095		
	7/7/15	Fox	MARK Johnson	.098		
	7/8/15	B. Connor	John Holleran Bugbee	.097		
	7/9/15	Metayer	Chris Perron	.098	.097	
	7/14/15	A. Collier	Mark Archambault	.095		
	7/21/15	B. Connor	Joseph T. Mikek	.098		

DATAMASTER DMT OPERATOR USE LOG

VSP Williston
Department/Agency

100132
DataMaster DMT Serial Number

JULY 2015

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7/26/15	M. LEBLANC, M	KASEY HASKINS	.095	.093
8/3/15	WARNER	MARILYN KNUTSON	.094	.095
8/6/15	K. DUCHARME N.P.D			

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

Williston VSP
Department/Agency

100132
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/4/15	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
6/12/15	RPC Performed / Simulator Solution change [For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 14-12-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/10/2014 Expires: 9/10/2015 OPENED: TDA/290	TODD Ambrose DMT Supervisor
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/03/2015	20:13:11	WILLISTON VSP	JACOB A METAYER	0.152		0.095		
07/07/2015	19:40:47	WILLISTON VSP	JAMES V FOX	0.078		0.096		
07/08/2015	00:26:24	WILLISTON VSP	BRIAN CONNOR	REFUSED		0.096		
07/08/2015	00:48:10	WILLISTON VSP	BRIAN CONNOR	0.044		0.097		
07/09/2015	22:58:41	WILLISTON VSP	JACOB A METAYER	REFUSED		0.096		
07/09/2015	23:19:32	WILLISTON VSP	JACOB A METAYER	0.116		0.097		
07/14/2015	01:17:51	WILLISTON VSP	ANDREW COLLIER	0.083		0.095		
07/21/2015	05:40:35	WILLISTON VSP	BRIAN CONNOR	0.225		0.096		
07/25/2015	23:57:52	WILLISTON VSP	MICHELLE A LEBLANC	0.101	0.105	0.095	0.095	

Status Record History

Status Date	Status Time	Test Type	Status Message