

# SOLUTION CHANGE

DataMaster DMT: 100134  
Location: Morristown PD  
Date: 05/21/2016  
Performed by: MICHAEL W REEVE



## Accuracy and Precision Check

Concentration = 0.098 g/210L  
Lot # = 16-04-100  
Range = 0.093 - 0.103  
Average = 0.100 g/210L  
Std Dev = 0.0005

*RCO*  
*5/23/16*

Simulator Temperature: 34.0°C

Performed by *Michael W Reeve*

Date 05/21/2016