Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message

Status Record History

Status Date	Status Time	Test Type	Status Message

DATAMASTER DMT MAINTENANCE LOG

FCSO
Department/Agency

100135
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/11	Set correct time [For RPC or Simulator Solution Change Place Label Here]	N. Vaillancourt
4/28	Jan, Feb. Mar Records To Lab	[illegible]
6/10	[For RPC **DataMaster Simulator Solution** Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 6/10/15 C. Miles Label Here]	[illegible]
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

FCSO
Department/Agency

100135
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
04/12/15	C. Miles	Heather Tynan	.098	.097
04/18/15	C. Miles	Peter Quebec	.097	
04/19/15	C. Miles	Justin Parah	.098	.098
05/22/15	N. Vaillancourt	Michael Hemingway	.096	
05/25/15	N. Vaillancourt	Seth Ovitt	.096	.095
06/2/15	Bushey-K.	Jeffrey Ketchum	.096	
06/07/15	N. Vaillancourt	Sheena Leveille	.096	
06/10/15	C. Miles	RPC	.099	
6/30/15	C. Miles	Josh Feeley	.099	.098

ALC 603 Rev 3 08/11