

SOLUTION CHANGE

DataMaster DMT: 100137
Location: Grand Isle County SO
Date: 01/28/2016
Performed by: BLAKE R ALLEN



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0000

ROD 1/29/16

Simulator Temperature: 34.1°C

Performed by *BLR*

Date 01/28/2016