

SOLUTION CHANGE

DataMaster DMT: 100137
Location: Grand Isle County SO
Date: 04/12/2016
Performed by: BLAKE R ALLEN



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 15-25-100
Range = 0.095 - 0.105
Average = 0.098 g/210L
Std Dev = 0.0005

RCD
4/13/16

Simulator Temperature: 34.1°C

Performed by

Blake R Allen

Date

04/12/2016