

# SOLUTION CHANGE

DataMaster DMT: 100137  
Location: Grand Isle County SO  
Date: 07/29/2016  
Performed by: BLAKE R ALLEN



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 16-05-100  
Range = 0.094 - 0.104  
Average = 0.098 g/210L  
Std Dev = 0.0005

*ROD 8/1/16*

Simulator Temperature: 34.0°C

Performed by *Blake R Allen*

Date 07/29/2016