

SOLUTION CHANGE

DataMaster DMT: 100138
Location: Derby VSP
Date: 01/04/2016
Performed by: GREGORY R MARSH



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-09-100
Range = 0.094 - 0.104
Average = 0.097 g/210L
Std Dev = 0.0004

ROD
1/5/16

Simulator Temperature: 34.0°C

Performed by

G. Marsh

Date

01/04/2016