

SOLUTION CHANGE

DataMaster DMT: 100140
Location: Bradford VSP
Date: 04/02/2016
Performed by: SEAN P BRENNAN



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0005

*RCD
4/4/16*

Simulator Temperature: 33.9°C

Performed by _____

Date 04/02/2016