BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

Vernon
Department/Agency

950159
DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
12/18/08	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 09-11-100 Conc: 0.103 Range at Installation: 0.098-0.108 Preparation Date: 10/14/08 Expires: 10/14/09 Installed Instrument	Steven Harnois
02/28/09	[Place Simulator Solution Label Here] Keyboard (internal) not working – display frozen removed cover – found keyboard plug loose checked connection – re-seated plug.	Sgt. Christopher Buckley VSP-Brattleboro
2/28/09	[Place Simulator Solution Label Here] RPC **DataMaster Simulator Solution** Lot # 09-18-100 $EtOH_{aq}$ Conc: 0.099 Range at Installation: 0.094-0.104 Preparation Date: 12/9/08 Expires: 12/9/09	Sgt. Christopher Buckley VSP-Brattleboro
06/26/09	[Place Simulator Solution Label Here] RPC **DataMaster Simulator Solution** Lot # 09-40-100 Conc: 0.100 Aq EtOH Range at Installation: 0.095-0.105 Preparation Date: 05/12/2009 Expires: 05/12/2010	Sgt. Christopher Buckley VSP-Brattleboro
06/26/09	[Place Simulator Solution Label Here] Corrected Time on DATAMASTER	Sgt. Christopher Buckley VSP-Brattleboro
11/02/09	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 10-15-100 Conc: 0.098 Aq EtOH Range at Installation: 0.093-0.103 Preparation Date: 10/13/2009 Expires: 10/13/2010	[illegible] Vernon P.D.

ALC 803 Rev. 2 06/07

BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

Vernon
Department/Agency

950159
DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
11/06/09	[Place Simulator Solution Label Here] RPC REDO	BRUCE GAULD VERNON P.D.
2/19/10	[Place Simulator Solution Label Here] RPC **DataMaster Simulator Solution** Lot # 10-15-100 Conc: 0.098 Aq EtOH Range at Installation: 0.093-0.103 Preparation Date: 10/13/2009 Expires: 10/13/2010	Bruce Gauld Vernon P.D.
6/28/10	[Place Simulator Solution Label Here] RPC **DataMaster Simulator Solution** Lot # 10-17-100 $EtOH_{aq}$ Conc: 0.103 Range at Installation: 0.098 – 0.108 Preparation Date: 11/30/2009 Expires: 11/30/2010	Bruce Gauld Vernon PD
10/19/10	[Place Simulator Solution Label Here] Removed For DMT Deployment	Steven Harnois
	[Place Simulator Solution Label Here]	
	[Place Simulator Solution Label Here]	

ALC 803 Rev. 2 06/07

BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

~~101708~~ Vernon Pd
Department/Agency

101708
DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
10/19/2010	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 11-04-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 8/19/2010 Expires: 8/19/2011 Install Per Procedure	Steven Harnois
10/29/2010	[Place Simulator Solution Label Here] Called for Simulator Time Out Could not Recreate error But Found Simulator Leaking, Repaired Leak and Tested w Accuracy and Precision test with	
	[Place Simulator Solution Label Here] no Problems Ran Subject Test with no Problems, Instrument Back in Service.	Steven Harnois
01/11/11	[Place Simulator Solution Label Here] Unit taken out of service and moved for office construction.	Bruce Gould
01/20/11	[Place Simulator Solution Label Here] Unit back in Service	Bruce Gould
01/20/11	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 11-04-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 8/19/2010 Expires: 8/19/2011 RPC	Bruce Gould

ALC 803 Rev. 2 06/07

BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

Vernon Police
Department/Agency

101708
DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
4/26/11	[Place Simulator Solution Label Here] options print out for Lab	Bruce Harold
6/16/11	RPC [Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 11-14-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 10/8/2010 Expires: 10/8/2011	Bruce Harold
8/17/11	[Place Simulator Solution Label Here] Removed Instrument from Site to VDHL	Steven Harnois
	[Place Simulator Solution Label Here]	
	[Place Simulator Solution Label Here]	
	[Place Simulator Solution Label Here]	

ALC 803 Rev. 2 06/07

DATAMASTER DMT MAINTENANCE LOG

Vernon PD
Department/Agency

100141
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-8-12	**DataMaster Simulator Solution** Lot # 12-51-100 $EtOH_{aq}$ Conc: 0.100 [For RPC or Simulator Solution Change Place Label Here] Range: 0.095 - 0.105 Preparation Date: 01/12/2012 Expires: 01/12/2013 OPENED: 2-8-12 Install Per Protocall	Steven Harnois K-K Kimball VDHC Chief
4/10/12	Replaced Simulator due to [For RPC or Simulator Solution Change Place Label Here] temperature issues. Performed A+P test on solution already in jar (see above) Temp checked by ALB + SMB	A. Bolo
6/23/12 RPC	[For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 12-51-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/12/2012 Expires: 01/12/2013 OPENED: 6/23/12	Bruce Gaul
8/22/12	Performed APM + Software update [For RPC **DataMaster Simulator Solution** ce Label Here] Lot # 12-64-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 OPENED: 8/22/12 RD Also had to do solution change protocol	R. Driscoll

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Vernon PD
Department/Agency

100141
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/5/12	RPC [For RPC ... ce Label Here] **DataMaster Simulator Solution** Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED:	Sgt. Bruce Gould
2/26/13	RPC [For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 OPENED:	Sgt Bruce Gould
6/12/13	**DataMaster Simulator Solution** Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 [For RPC ... e Label Here] OPENED: 6/7/13 Performed RPC and APM protocols.	R. Driscoll
10/30/13	**DataMaster Simulator Solution** Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 [For RPC ... ce Label Here] OPENED: 10/30/13	B. Gould

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Vernon Police
Department/Agency

100141
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/25/14	[For RPC **DataMaster Simulator Solution** RPC Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 ce Label Here] Preparation Date: 05/10/2013 Expires: 05/10/2014 OPENED: 2/25/14	Bruce Stanfield
6/3/14	[For RPC **DataMaster Simulator Solution** Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Label Here] Preparation Date: 3/12/2014 Expires: 3/12/2015 OPENED: ~~[illegible] 6/3/14~~ RPC	R Driscoll DMT TECH
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011