

SOLUTION CHANGE

DataMaster DMT: 100141
Location: Windham County SO - Vernon
Date: 08/05/2015
Performed by: ROBERT DRISCOLL



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-09-100
Range = 0.094 - 0.104
Average = 0.096 g/210L
Std Dev = 0.0004

Simulator Temperature: 33.9°C

Performed by RD

Date 08/05/2015

SOLUTION CHANGE

DataMaster DMT: 100141
Location: Windham County SO - Vernon
Date: 08/05/2015
Performed by: ROBERT DRISCOLL



Accuracy and Precision Check

Concentration
Lot #
Range
Average
Std Dev

Simulator Temperature:

Solution Change Failed

Simulator timed out. Rerun solution change protocol. RCD 8/5/15

Performed by

RCD

Date

08/05/2015