

SOLUTION CHANGE

DataMaster DMT: 100141
Location: WINDHAM COUNTY SO - VERNON
Date: 05/13/2016
Performed by: CHRIS NORTON



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.099 g/210L
Std Dev = 0.0005

*RD
5/16/16*

Simulator Temperature: 34.0°C

Performed by

CHRIS NORTON

Date

05/13/2016

SOLUTION CHANGE

DataMaster DMT: 100141
Location: WINDHAM COUNTY SO - VERNON
Date: 04/26/2016
Performed by: CHRIS NORTON



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-05-100
Range = 0.094 - 0.104
Average = 0.099 g/210L
Std Dev = 0.0004

Simulator Temperature: 34.0°C

RCD 4/26/16
Expired lot. Contacted
agency to remove from
service until valid
solution change completed.
Verified no breath tests
have been run on the instrument
since the solution change using
DMT host.

Performed by

CRN

Date

04/26/2016