DATAMASTER DMT MAINTENANCE LOG

BRANDON PD
Department/Agency

100142
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/7/16	OPENED: **DataMaster Simulator Solution** [For RPC Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 1/7/16 OPENED: e Label Here] SIM SOLUTION CHANGE	D. BUTTERFIELD DMT SUPERVISOR
2/20/16	**DataMaster External Standard Solution** Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 [For Preparation Date: 11/12/2015 Expires: 11/12/2016 OPENED: 2/20/16 Place Label Here] RPC	D. BUTTERFIELD DMT SUPERVISOR
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

Brandon P.D.
Department/Agency

100142
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
9-16-15	Michael Winkler	Ellis S. Grace II	.099	
11-13-15	BENJAMIN HERRICK	TIMOTHY J. DUFFY	.097	
12/3/15	R. Driscoll	TEST / APM	.095	—
1/6/16	Michael Winkler	Howard N. Lynk	.092	
1/6/16	Michael Winkler	Howard N. Lynk	.092	
1/6/16	Michael Winkler	Howard N. Lynk	.091	
1/7/16	D. BUTTERFIELD	SOLUTION CHANGE	.097	
2/14/16	D. BUTTERFIELD	WILLIAM E WOODARD	.096	
2/17/16	D. BUTTERFIELD	ROBERT C. WIGGIN	.096	
2/20/16	D. BUTTERFIELD	RPC	.100	