

# SOLUTION CHANGE

DataMaster DMT: 100142  
Location: Brandon PD  
Date: 08/04/2016  
Performed by: DAVID B BUTTERFIELD



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 16-03-100  
Range = 0.094 - 0.104  
Average = 0.100 g/210L  
Std Dev = 0.0005

RCD  
8/5/16

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "D. Butterfield", written over a horizontal line.

Date

08/04/2016