DATAMASTER DMT MAINTENANCE LOG

Randolph PD
Department/Agency

100143
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/9/15	**DataMaster Simulator Solution** Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 [For RF ... ace Label Here] OPENED: 10/09/15 Test / RPC	Steven Tiersch Officer DMT Supervisor
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

RANDOLPH P.D.
Department/Agency

100143
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
7/24/15	Tom Simpson	Ernest Cassella	.098	—
7/25	D. Frattini	BLANK ERROR HORNE, ALFRED	.195	—
7/25	D. Frattini	HORNE, ALFRED	.161	—
8/1/15	S. Lambert	Donahue, Thomas	.095	—
8/14/15	C. Garonzie	Ballou, Christopher	.099	—
9/1/15	Christopher Quesnel	Tyler Laperle	.098	
9/1/15	Christopher Quesnel	Tyler Laperle	.096	
9/13/15	Steven Tiersch	Beriah Smith	.097	
9/13/15	Steven Tiersch	Beriah Smith	.097	
10/09/15	Steven Tiersch	Test/RPC	0.098	

ALC 603 Rev 3 08/11