



VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: Brighton PD DataMaster Serial #: 100145

Date: 3/5/2016 Requestor: Jeff Noyes

Instrument Supervisor: Jeff Noyes Phone #: (802)723-4322

Taken By: Robert Driscoll

Problem: Simulator Paddle stopped spinning.

Laboratory Response: Alcohol program staff will troubleshoot onsite.

Work Performed: Replaced the simulator with Guth S/N: G4860. Matched temperatures using NIST traceable thermometer S/N: 111814367. Temperature reading 34.0 degrees C.

On Site In-house

Attachments: 2 pages

Performed By: Robert Driscoll
Final Review: [Signature]

Date: 3/8/2016
Date: 3/10/16

Page 1 of 3 KCS

Driscoll, Rob

From: Noyes, Jeffrey
Sent: Monday, March 07, 2016 8:50 AM
To: Driscoll, Rob
Subject: RE: Brighton DMT

Perfect. I will also need solution.

Jeffery M. Noyes, Chief Brighton Police Department 23 East Brighton Road PO Box 189 Island Pond VT 05846 802-723-4322 802-723-4486 fax Jeffrey.Noyes@vermont.gov Please note new e-mail address	Canaan Police Department 318 Christian Hill Road PO Box 159 Canaan VT 05903 802-266-9619 802-266-8253 fax
--	--

From: Driscoll, Rob
Sent: Monday, March 07, 2016 8:49 AM
To: Noyes, Jeffrey
Subject: RE: Brighton DMT

I can be there around noon if that works for you?

From: Noyes, Jeffrey
Sent: Monday, March 07, 2016 8:47 AM
To: Driscoll, Rob
Subject: RE: Brighton DMT

I will be here if you are coming.

Jeffery M. Noyes, Chief Brighton Police Department 23 East Brighton Road PO Box 189 Island Pond VT 05846 802-723-4322 802-723-4486 fax Jeffrey.Noyes@vermont.gov Please note new e-mail address	Canaan Police Department 318 Christian Hill Road PO Box 159 Canaan VT 05903 802-266-9619 802-266-8253 fax
--	--

From: Driscoll, Rob
Sent: Monday, March 07, 2016 8:41 AM
To: Noyes, Jeffrey
Subject: RE: Brighton DMT

Jeff,

Are you going to be in Brighton today? I can bring a new simulator.

Regards,
Rob

From: Noyes, Jeffrey
Sent: Saturday, March 05, 2016 1:45 PM
To: Driscoll, Rob
Subject: Brighton DMT

Rob,

I went to do the RPC on Brighton's machine today and ran into a little snag. When I plugged the power cord back in the paddle wasn't rotating, so as instructed I pulled the paddles out of the solution and lightly pushed the paddles to get them rotating. They started to rotate then smoke came out of the mechanism head. I immediately unplugged the machine and have placed it out of service.

Jeffery M. Noyes, Chief
Brighton Police Department
23 East Brighton Road
PO Box 189
Island Pond VT 05846
802-723-4322
802-723-4486 fax
Jeffrey.Noyes@vermont.gov
Please note new e-mail address

Canaan Police Department
318 Christian Hill Road
PO Box 159
Canaan VT 05903
802-266-9619
802-266-8253 fax