ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100148
Location: Berlin PD
Calibration Date: 11/09/2011
Certification Date: 11/14/2011
Installation Date: 12/08/2011
RPC Date: 06/07/2012
Supervisor Name: CHAD J BASSETTE

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.06
Modem: 2.2
Questions: 2.0

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube    = 48.1°C
Digital Sim    = 34.0°C

SETTINGS
Lamp Voltage   = 1.64 V
Cooler Voltage = 1.69 V
Bias Voltage   = 80 V
Chopper Freq   = 538 Hz

PUMP INFO
Flow Rate = 5.689 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.1098    0.1128
MIN(V)   0.1081    0.1110

FILTER INFO
Filter 1   0.111  Zero = true
Filter 2   0.464  Zero = true
Filter 3   0.331  Zero = true

CALIBRATION CHECK
Xq  = 0.083  0.10%
```

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.101 g/210L
Lot #	= 12-64-100
Range	= 0.096 - 0.106
Average	= 0.101 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by _____ **Date** 06/07/2012

Reviewed by _____ **Date** 6/11/12

6/12/2012