DATAMASTER DMT OPERATOR USE LOG

Berlin
Department/Agency

100148
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/30/14	Bullard, Jonathan	Andrew Barr	.100	.099
9/15/14	Tyler/Bassette, Chad	Tyler/Eckerson	.096	
9-20-14	Kolpitarski, Kyle	Justin Galbais	.098	
10/25/14	RPC	RPC	.097	
10/27/14	Quesnel, Chris	Robert Richardson	.097	
11/2/14	Bullard, Jonathan	James Thompson	.097	
11/12/14	R. Driscoll	TEST / APM	.098	—
11/15/14	Bullard, Jonathan	Thomas Worden	.097	.096
11/22/14	Bullard, Jonathan	Adrianna Milo	.097	.096
12-19-14	Montelth, Mark	Robert Ambrown	.095	.095
2/7/14	Bassette, Chad	RPC	.098	

Log On Record History

Test Date	Test Time	Name

Status Record History

Status Date	Status Time	Test Type	Status Message

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message

DATAMASTER DMT MAINTENANCE LOG

Berlin PD
Department/Agency

100148
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/12/14	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.
2/7/14	[For RP **DataMaster Simulator Solution** Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: ~~7/14/2015~~ OPENED: CS 2/7/15 ce Label Here]	CS
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011