

SOLUTION CHANGE

DataMaster DMT: 100148
Location: Berlin PD
Date: 11/29/2015
Performed by: CHAD J BASSETTE



Accuracy and Precision Check

Concentration = 0.097 g/210L
Lot # = 15-23-100
Range = 0.092 - 0.102
Average = 0.099 g/210L
Std Dev = 0.0005

RP
11/30/15

Simulator Temperature: 34.0°C

Performed by

Date

11/29/2015