

SOLUTION CHANGE

DataMaster DMT: 100149
Location: Barre City PD
Date: 09/03/2015
Performed by: ROBERT DRISCOLL



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-12-100
Range = 0.094 - 0.104
Average = 0.096 g/210L
Std Dev = 0.0004

Simulator Temperature: 33.9°C

Performed by

RD

Date

09/03/2015