

SOLUTION CHANGE

DataMaster DMT: 100149
Location: Barre City PD
Date: 12/26/2015
Performed by: JUSTIN L PICKEL



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.096 g/210L
Std Dev = 0.0005

*ROD
12/28/15*

Simulator Temperature: 33.9°C

Performed by



Date

12/26/2015
