

# SOLUTION CHANGE

DataMaster DMT: 100149  
Location: Barre City PD  
Date: 04/05/2016  
Performed by: TROY S KIRKPATRICK



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 16-03-100  
Range = 0.094 - 0.104  
Average = 0.096 g/210L  
Std Dev = 0.0004

*RCD*  
*4/6/16*

Simulator Temperature: 34.0°C

Performed by

*T. S. Kirkpatrick*

Date

04/05/2016