ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100152
Location: Woodstock PD
Calibration Date: 03/07/2012
Certification Date: 03/07/2012
Installation Date: 03/08/2012
RPC Date: 05/30/2014
Supervisor Name: JAMES P OTRANTO

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.06
Modem: 2.4
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube    = 48.1°C
Digital Sim    = 34.0°C

SETTINGS
Lamp Voltage   = 1.71 V
Cooler Voltage = 1.73 V
Bias Voltage   = 81 V
Chopper Freq   = 554 Hz

PUMP INFO
Flow Rate = 5.680 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.5931    0.5959
MIN(V)   0.5911    0.5925

FILTER INFO
Filter 1   0.595   Zero = true
Filter 2   0.889   Zero = true
Filter 3   1.463   Zero = true

CALIBRATION CHECK
Xq  = 0.080  1.32%
```

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.099 g/210L
Lot #	= 13-21-100
Range	= 0.094 - 0.104
Average	= 0.097 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by James P Otranto **Date** 05/30/2014

Reviewed by _____ **Date** 6/5/14

6/5/2014