Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/04/2015	04:18:53	BRATTLEBORO PD	ADAM S BELVILLE	0.162		0.098	0.096	Blank Error
04/04/2015	04:28:16	BRATTLEBORO PD	ADAM S BELVILLE	0.155		0.098		
04/09/2015	01:10:37	WINDHAM COUNTY SO	IAN A TUTTLE	0.144	0.142	0.097	0.097	
04/24/2015	01:53:37	BRATTLEBORO PD	CARL W WARNER	REFUSED		0.096		

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/23/15	Screen frozen - turned power off [For RPC or Simulator Solution Change Place Label Here] and restarted	P. Withubee M1239
2-1-15	[For RP ... ace Label Here] DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 2-1-15 AB M1219	SGT. Adam Belville AdBelville M1219
Sent to SA 2-2-15	RPC - Adam Belville 2-1-15	
3-11-15	Adjusted time for Daylight Savings [For RPC or Simulator Solution Change Place Label Here]	SGT Adam Belville ABelville M1219
	[For RPC or Simulator Solution Change Place Label Here]	

Status Record History

Status Date	Status Time	Test Type	Status Message
04/04/2015	04:25:58	DUI Subject Record	Blank Error