DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-3-16	[For RPC or Simulator Solution Change Place Label Here] RPC **DataMaster External Standard Solution** Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: 6-3-16 AB MJD19	SGT. Adam Belville MJD19
8/2/16	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
8-12-16	Ran Accuracy : Precision Report [For RPC or Simulator Solution Change Place Label Here] DMT stated Solution out of Range During DUI processing	SGT. Adam Belville MJD19
8/22/16	**DMT External Standard Solution** Lot # 16-15-100 EtOHaq Conc: 0.100 [For RPC or S Range: 0.095 - 0.105 ...el Here] Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 8/22/16 RCD changed solution	R. Driscoll DMT Tech

ALC 803 REV 03 08/2011