DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/22/15	[For RPC **DataMaster Simulator Solution** Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 OPENED: 7/22/15 RCD e Label Here] Installation post TSI	R. Driscoll DMT Tech.
10-02-15 Sent to SA 10-02-15	[For RPC **DataMaster Simulator Solution** Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 10-02-2015 AB MNI9 e Label Here] RPC - Adam Belville 10-02-15	SGT. Adam Belville ABelville MNI9
10/23/15	[For RPC **DataMaster Simulator Solution** Lot # 15-13-100, EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 10/23/15 RCD e Label Here] Install post TSI dated 10/22/15	R. Driscoll DMT Tech.
11-1-15	Adjusted for Daylight Savings [For RPC or Simulator Solution Change Place Label Here]	SGT. Adam Belville ABelville MNI9

ALC 803 REV 03 08/2011