DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED B (Name and Title)
2-1-16	Re-Started DataMaster : Printer [For RPC or Simulator Solution Change Place Label Here] DataMaster wasn't communicating with printer 2-1-16 @ 21:44 hrs	SGT. Adam Belville MIDI9
2-1-16	[For RPC or Simulator Solution Change Place Label Here] RPC **DataMaster Simulator Solution** Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 2-1-16 AB MIDI9	SGT. Adam Belville MIDI9
5-19-16	Re-Started DataMaster. [For RPC or Simulator Solution Change Place Label Here] DataMaster wasn't communicating with the printer. 5-19-16 @ 22:01 hrs	SGT. Adam Belville MIDI9
5-20-16	Re-Started DataMaster [For RPC or Simulator Solution Change Place Label Here] "out of range" Timer restarted for DUI then remained "busy" for an extended period of time. 5-20-16 @ 03:30 hrs	SGT. Adam Belville MIDI9

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-3-16	[For RPC or Simulator Solution Change Place Label Here] RPC **DataMaster External Standard Solution** Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: 6-3-16 AB MJ19	SGT. Adam Belville MJ19
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011