DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD
Department/Agency

100153
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-1-16	Re-started DataMaster & Printer [For RPC or Simulator Solution Change Place Label Here] DataMaster wasn't communicating with printer 2-1-16 @ 21:44 hrs	SGT. Adam Belville MIDI9
2-1-16	[For RPC or Simulator Solution Change Place Label Here] RPC **DataMaster Simulator Solution** Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 2-1-16 AB MIDI9	SGT. Adam Belville MIDI9
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011