

SOLUTION CHANGE

DataMaster DMT: 100153
Location: BRATTLEBORO PD
Date: 08/22/2016
Performed by: ROBERT DRISCOLL



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 16-15-100
Range = 0.095 - 0.105
Average = 0.100 g/210L
Std Dev = 0.0004

Simulator Temperature: 34.0°C

Performed by

RD

Date

08/22/2016