

# SOLUTION CHANGE

DataMaster DMT: 100159  
Location: Milton PD  
Date: 12/31/2015  
Performed by: ED A LARENTE



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-12-100  
Range = 0.094 - 0.104  
Average = 0.098 g/210L  
Std Dev = 0.0004

ROD  
1/4/16

Simulator Temperature: 34.1°C

Performed by

Ed A Larente

Date

12/31/2015