2012 DataMaster DMT APM and Software Update

Date	5/31/12	
Agency	SBPD	
DMT Serial Number	100162	
Technician Name	[illegible]	
Software Updated	YES	NO
Lock Changed	YES	NO
DPS Property Tag Number	20165	
APM Protocol Performed	YES	NO
Other/Notes	Touch screen failed ~~Out of Svc.~~ left in service. can still be used w/out the touch feature [illegible] 5/31/12	