DATAMASTER DMT OPERATOR USE LOG

So. Burlington
Department/Agency

[illegible]
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/19	DeFiore	Colby Martin	.097	.095
8/20	Matott	Marlene Trembach	.097	
8/20	Pope	Todd Ellsworth	.096	
8/20	Matott	Daniel Fischer	.096	.095
8/22	Demore	Ramon Burgos	.100	
8/22	Pope	Alyssa Bennett	.096	
8/24	Brunelle	Nicholas Cota	.096	
8/27	B. Williamson	Maria Gould	.095	out of range .094
8/28	Mills	TEST/Sim Sol Change	.095	
9/4	DeFiore	Barnah Koreyen	out of range .094	0.94
9/7	Matott	Travis Barewicz	out of range .094	.093
9/8	Superneau	Solution change	.097	
9/18	Brunelle	Robert Petherick	.098	.094
9/20	Matott	Jennifer Mitchell	.098	
9/23	WILLIAMSON	CALEB FONTNEAU	.096	
9/23	Crispin	Ren Chase	.097	.095
10/2/15	R. Driscoll	TEST / APM	.097	
10/2/15	R Driscoll	TEST / check-in	.096	—
10/4/15	BOYER	KAYNE BURDO	.097	
10/8/15	Pope	Thomas Burrows	.096	

ALC 603 Rev 3 08/11

DATAMASTER DMT OPERATOR USE LOG

So. Burlington
Department/Agency

100162
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
10/8	Pope	Thomas Burrows	.096	
10/9/15	October 2015 RPC	DET. Superneau	.099	
10/10/15	Matott	Travis Babcock	.099	
10/11/15	Mulcahy	Joshua Horton	.099	.098
10/16/15	Pope	Zachary Whitehouse	.099	
10/17/15	Pope	Crystal Santana	.098	
10/25/15	Mulcahy	Bernardo Nieves	.099	.098
10/26/15	Pope	Jean Shungu	.098	.096
10/30/15	Pope	Jean Shungu	.097	.097
10/31/15	Bataille	Tayla Vanasse	.099	
10/31/15	Pope	Jon Cabral	.097	
11/3/15	Pope	James Grace	.098	
11/3/15	Pope	James Grace	.098	
11/5/15	Matott	Christopher Manning	.097	.097
11/8/15	Pope	Matthew Logan	.097	.098
11/8/15	Pope	Cian McGrine	.096	.096

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

SoBurlington DM
Department/Agency

100162
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/2/15	Completed APM protocol + Downloaded data [For RPC or Simulator Solution Change Place Label Here] to usb drive, uploaded new software, installed new I/O PCB. Port not active at this time	R. Driscoll DMT Tech.
10/9/2015	[For RPC **DataMaster Simulator Solution** Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 e Label Here] OPENED: 10/9/15 OCTOBER RPC 2015	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011