

# SOLUTION CHANGE

DataMaster DMT: 100162  
Location: SOUTH BURLINGTON PD  
Date: 04/04/2016  
Performed by: SARAH SUPERNEAU



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-24-100  
Range = 0.094 - 0.104  
Average = 0.098 g/210L  
Std Dev = 0.0000

*RCD  
4/5/16*

Simulator Temperature: 33.9°C

Performed by

*Sarah Superneau 654*

Date

04/04/2016