

SOLUTION CHANGE

DataMaster DMT: 100162
Location: SOUTH BURLINGTON PD
Date: 05/02/2016
Performed by: SARAH SUPERNEAU



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 16-03-100
Range = 0.094 - 0.104
Average = 0.097 g/210L
Std Dev = 0.0004

RCD
5/3/16

Simulator Temperature: 33.9°C

Performed by

654

Date

05/02/2016