

SOLUTION CHANGE

DataMaster DMT: 100163
Location: Bennington County SO
Date: 08/24/2015
Performed by: ROBERT DRISCOLL



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 15-06-100
Range = 0.095 - 0.105
Average = 0.098 g/210L
Std Dev = 0.0004

Simulator Temperature: 33.7°C

Performed by

RD

Date

08/24/2015

RD
8/25/15