

SOLUTION CHANGE

DataMaster DMT: 100164
Location: Rutland City PD
Date: 01/07/2016
Performed by: CHARLES G WHITEHEAD



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-05-100
Range = 0.094 - 0.104
Average = 0.096 g/210L
Std Dev = 0.0004

ROD
1/8/16

Simulator Temperature: 34.0°C

Performed by

C. Whitehead

Date

01/07/2016