SOLUTION CHANGE

DataMaster DMT: 100164
Location: Rutland City PD
Date: 01/27/2016
Performed by: CHARLES G WHITEHEAD

Accuracy and Precision Check

Concentration	= 0.099 g/210L
Lot #	= 15-24-100
Range	= 0.094 - 0.104
Average	= 0.099 g/210L
Std Dev	= 0.0000

RCD 1/28/16

Simulator Temperature: 34.0°C

Performed by _____ Date 01/27/2016