2012 DataMaster DMT APM and Software Update

Date	6/8/12	
Agency	VSP- Shaftsbury	
DMT Serial Number	100166	
Technician Name	AJB	
Software Updated	YES	NO
Lock Changed	YES	NO
DPS Property Tag Number	20169	
APM Protocol Performed	YES	NO
Other/Notes		

TSI -
Replaced simulator w/ new. New sim did not have S/N sticker so I labeled it "VT11" Performed pressure check and adjusted sim temp to read 34.1°C using thermometer S/N 111814522. Performed Installation protocol to satisfy both the RPC and the APM

6/8/12
A. Bolden