



National Patent Analytical Systems, Inc.

On this date 10/7/11, NPAS received instrument # 100167.

The instrument was sent from the Vermont Department of Health (VDH) so that NPAS could install updated software as well as inspect the instrument. At NPAS' request VDH provided copies of the Technical Support Inquiries (TSI) regarding the instrument. The calibration history of the instrument remained in the instrument so that NPAS could review its history.

NPAS examined the internal configuration of the instrument and did / did not find physical manipulation or tampering of the components located within the instrument itself.

NPAS did performance and accuracy checks on a certified 0.100 ethanol standard. NPAS found that the instrument passed / failed the performance and accuracy checks.

NPAS reviewed the instrument as to its DataMaster DMT Inspection, QA and Conformance Testing Procedure and the instrument performed / did not perform within its procedure.

NPAS examined the calibration records and did / did not observe anomalies.

NPAS testing indicates that there is no reason why this instrument would not have performed accurately and reliably in the field.

Print Name of Evaluator: Michael J. Barbely Jr.

Signature of Evaluator and Date: Michael J. Barbely Jr. 10/10/11

Print Name of Reviewer: DAVID RADOMSKI

Signature of Reviewer and Date: David Radomski 10/10/11

*opened  
10-13-11  
MMA*  
*[Signature]*  
10/13/11

**Packing List**

From: National Patent Analytical Systems Inc.  
2090 Harrington Mem. Rd., Mansfield, Ohio 44903

Cust No. VT1678J Ord Date 10/7/11  
Sales Order No. 44841

Ship Date 10/11/11  
Packing List No. 69071

Cust Purch Ord No. KIRK

Ship To:  
Vermont Dept of Health

195 Colchester Avenue  
Burlington Vt 05401

Bill To:  
Vermont Dept of Health

P O Box 70  
Burlington Vt 05402

25.00

Freight

Attn: Ship Via FedEx

Comments Service for Instrument Serial Number 100167

Inv Part No	Quantity Ordered	This Shipment	Back Ordered	Total Ship To Date	Description	Customer Part No.	Drawing Part No. / Rev
9DMT	1.0	1		1.0	DATAMASTER DMT SERVICE RETURN	None Given	No Draw # / NONE
S/N 100167							
Serial Nos.							

*Handwritten:*  
10-13-11  
Jan 10/13/11