

# SOLUTION CHANGE

DataMaster DMT: 100168  
Location: WINHALL PD  
Date: 11/25/2015  
Performed by: JEREMIAH J ROGERS



## Accuracy and Precision Check

Concentration = 0.099 g/210L  
Lot # = 15-13-100  
Range = 0.094 - 0.104  
Average = 0.100 g/210L  
Std Dev = 0.0004

*ROD 11/25/15*

Simulator Temperature: 33.9°C

Performed by

*Jeremiah J Rogers*

Date

11/25/2015

# SOLUTION CHANGE

DataMaster DMT: 100168  
Location: WINHALL PD  
Date: 11/25/2015  
Performed by: JEREMIAH J ROGERS



## Accuracy and Precision Check

Concentration  
Lot #  
Range  
Average  
Std Dev

Simulator Temperature:

Solution Change Failed

Performed by

A handwritten signature in black ink, appearing to read "Jeremiah J. Rogers".

Date

11/25/2015