

## SOLUTION CHANGE

DataMaster DMT: 100169  
Location: Bellows Falls PD  
Date: 01/28/2015  
Performed by: MICHAEL J KEEFE



### Accuracy and Precision Check

Concentration = 0.097 g/210L  
Lot # = 14-10-100  
Range = 0.092 - 0.102  
Average = 0.101 g/210L  
Std Dev = 0.0000

Simulator Temperature: 34.1°C

Performed by

A handwritten signature in black ink, appearing to read "Michael J. Keefe".

Date

01/28/2015