SOLUTION CHANGE

DataMaster DMT: 100171
Location: St. Albans VSP
Date: 01/06/2016
Performed by: PATTY A WHITNEY

Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-07-100
Range	= 0.095 - 0.105
Average	= 0.097 g/210L
Std Dev	= 0.0000

RD 1/7/16

Simulator Temperature: 34.2°C

Performed by Patrina A Whitney Date 01/06/2016