

SOLUTION CHANGE

DataMaster DMT: 100171
Location: St. Albans VSP
Date: 10/02/2016
Performed by: RIGGEN JAY



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.098 g/210L
Std Dev = 0.0004

RCD 10/3/16

Simulator Temperature: 34.1°C

Performed by



Date

10/02/2016
