

SOLUTION CHANGE

DataMaster DMT: 101708
Location: Royalton VSP
Date: 04/11/2016
Performed by: STEVEN C STEARNS



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 15-07-100
Range = 0.095 - 0.105
Average = 0.101 g/210L
Std Dev = 0.0004

ROD 4/11/16

Simulator Temperature: 34.1°C

Performed by

Steven Stearns

Date

04/11/2016