SOLUTION CHANGE

DataMaster DMT: 102506
Location: Lamoille County SO
Date: 04/02/2016
Performed by: CHRISTOPHER M JONES

Accuracy and Precision Check

Concentration	= 0.100 g/210L
Lot #	= 15-25-100
Range	= 0.095 - 0.105
Average	= 0.097 g/210L
Std Dev	= 0.0004

RCD 4/4/16

Simulator Temperature: 34.2°C

Performed by [signature] Date 04/02/2016