DataMaster DMT Monthly Update Check-list

Department/Agency: Stowe Police DM serial #: 102706

Supervisor: Knight Date: 11 14 15

	Provided	None this month
Operators Logbook	☑	☐
Maintenance Logbook	☑	☐
Breath Summary Report	☑	☐
Status Summary Report	☑	☐
Calibration Report	☐	☐
Certification Report	☐	☐
Installation Report	☐	☐
Annual Preventative Maintenance	☐	☐
Routine Performance Check	☐	☐
Simulator Solution Change	☐	☐
Other Reports	☐	☐

List other report type and date:

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
10/08/2015	00:38:21	STOWE PD	DAVID M KNIGHT	0.124	0.125	0.094	0.094	
10/10/2015	03:25:01	STOWE PD	R CASHIN	REFUSED		0.096		
10/10/2015	03:46:41	STOWE PD	RYAN CASHIN	0.167	0.174	0.096	0.095	
10/24/2015	03:36:00	STOWE PD	RYAN CASHIN	0.178		0.094		RESPONSE TIMEOUT

Status Record History

Status Date	Status Time	Test Type	Status Message
10/24/2015	03:40:27	DUI Subject Record	RESPONSE TIMEOUT

DATAMASTER DMT OPERATOR USE LOG

Stowe Pd
Department/Agency

102706
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
1/26/12	Steven Harnois	Install	.100	
1/29/12	Sean Dulina			
1/10/13	R. Driscoll	TEST / APM	.095	—
1/13/14	R. Driscoll	TEST / APM	.095	—
4-14-4	D. Knight	Test - Sim Sol change	.096	
5/3/14	D. Knight	Sol change	.096	
5/25/14	D. Knight	RPC - June	.095	
7/4/14	D. Knight	Test / simulator Sol Change	.096	
9-11-14	D. Knight	Test / Solution change	.096	
10-8-14	D. Knight	RPC - October	.097	
1-9-15	D. Knight	Solution change	.095	
1/15/15	R. Driscoll	TEST / APM	.094	
1-15-15	D. Knight	Solution change	.095	
2-6-15	D. Knight	RPC - February	.096	
2-13-15	K. Walker	TEST Solution Change	.095	
5-23-15	D. Knight	Solution Change	.096	
6-13-15	D. Knight	RPC - June	.097	
7-31-15	D. Knight	Solution change	.097	
9/30/15	R. Driscoll	TEST / check-in	.094	—
10/8/15	D. Knight	RPC - October	.097	—

1-26-12

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

Stowe PD
Department/Agency

102706
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/30/15	Downloaded data to USB drive. Uploaded [For RPC or Simulator Solution Change Place Label Here] new software, and installed new I/O PCB. No port available at this time	R. Driscoll DMT Tech.
10/5/15	OPENED: **DataMaster Simulator Solution** [For RF Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 ace Label Here] OPENED: October RPC	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	