Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim #G11309 NIST Temp = 34.1°C 3-19-12 Sim 03/19/12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 3-19-12

Re:
Shipment of DataMaster DMT S/N: 102806 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 3-19-12
2. **Removed breath tube.** 3-19-12
3. **Downloaded instrument files to USB thumb drive.** 3-19-12
4. **Removed keyboard & stylus.**
5. **Removed security screws & replaced with manufacturer supplied screws.** No reg. screws available sent w/ security screws 3-19-12
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 8

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: 3-19-12 **Witness by:** Sim 03/19/12

Franklin County S.O. St Albans.

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/28/2010	11:13	GBU	YUGBUI G BGUI			0.097		USER ABORT
09/28/2010	11:15	GBU	YUGBUI G BGUI	0.100		0.099		
09/28/2010	11:22	JIOP	GYU9O G FGYUI	0.000		0.000		
09/28/2010	11:30	FTI	FTYUI D FT7I	0.384		0.000		
10/09/2010	01:29	ST. ALBANS PD	JASON WILLIAMS	0.122	0.109	0.097	0.096	
11/19/2010	01:58	ST. ALBANS PD	SEAN E POPE	0.079		0.096		
11/20/2010	18:05	FRANKLIN COUNTY SO	EDOUARD A LARENTE	0.163		0.095		
11/21/2010	04:02	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE	0.123		0.096		
11/25/2010	03:44	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE			0.096		RESPONSE TIMEOUT
11/25/2010	03:49	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE	R		0.096		
11/25/2010	04:09	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE	0.179	0.188	0.094	0.093	
11/28/2010	03:25	FRANKLIN COUNTY SO	SHAUN M SENECAL	0.059		0.095		
12/12/2010	00:30	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.124		0.095		
01/21/2011	03:01	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			0.090		STANDARD OUT OF RANGE
01/21/2011	03:19	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			0.090		STANDARD OUT OF RANGE
02/17/2011	21:46	FRANKLIN COUNTY SO	CHAD A MILES	0.154		0.097	0.095	STANDARD OUT OF RANGE
02/26/2011	22:16	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.083	V	0.101	0.100	
02/26/2011	22:42	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.079		0.100		
02/27/2011	03:41	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.138	0.140	0.100	0.099	
03/12/2011	04:19	ST. ALBANS PD	SEAN E POPE	0.085		0.101	0.101	DETECTOR OVERFLOW
03/12/2011	04:26	ST. ALBANS PD	SEAN E POPE					FILTER 3 WON'T ZERO
03/12/2011	04:29	ST. ALBANS PD	SEAN E POPE					FILTER 3 WON'T ZERO
03/13/2011	03:07	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD					DETECTOR OVERFLOW
03/13/2011	03:09	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD					DETECTOR OVERFLOW
03/27/2011	21:25	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.254	0.264	0.100	0.099	
04/04/2011	21:05	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.097	0.099	0.099	0.098	
04/08/2011	19:16	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.177		0.099		
04/09/2011	21:04	FRANKLIN COUNTY SO	ERIC W THOMPSON	0.182	0.164	0.099	0.097	
04/10/2011	03:12	FRANKLIN COUNTY SO	JARED J MITCHELL	0.101	0.098	0.099	0.099	
04/22/2011	20:00	FRANKLIN COUNTY SO	BENJAMIN M PATNODE	0.080		0.099		
04/25/2011	01:44	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.130	0.130	0.099	0.099	
05/04/2011	22:59	FRANKLIN COUNTY SO	BRENDAN J MCKENNEY	0.164	0.167	0.098	0.098	
05/05/2011	02:16	FRANKLIN COUNTY SO	JEROD J ADAMS	0.130		0.098		
05/06/2011	20:23	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE			0.099		RESPONSE TIMEOUT
05/06/2011	20:46	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE	0.095	0.098	0.099	0.098	
05/08/2011	00:43	FRANKLIN COUNTY SO	W RUPRECHT	0.186	0.183	0.098	0.097	
05/08/2011	02:25	FRANKLIN COUNTY SO	DEAN M SCOTT			0.091		STANDARD OUT OF RANGE
05/08/2011	02:28	FRANKLIN COUNTY SO	DEAN M SCOTT			0.093		STANDARD OUT OF RANGE
05/11/2011	21:55	FRANKLIN COUNTY SO	JACY L DENNETT	0.096	0.096	0.101	0.100	
05/21/2011	18:43	FRANKLIN COUNTY SO	CHAD A MILES	R		0.095		
05/28/2011	03:03	FRANKLIN COUNTY SO	JARED J MITCHELL	R		0.097		
05/28/2011	03:27	FRANKLIN COUNTY SO	JARED J MITCHELL			0.094		STANDARD OUT OF RANGE
05/28/2011	03:30	FRANKLIN COUNTY SO	JARED J MITCHELL	0.154		0.095		
05/29/2011	18:04	FRANKLIN COUNTY SO	BRIAN M BEGNOCHE	0.125		0.095	0.094	STANDARD OUT OF RANGE
06/25/2011	01:26	FRANKLIN COUNTY SO	JARED J MITCHELL	0.131		0.100		
06/26/2011	05:41	FRANKLIN COUNTY SO	JARED J MITCHELL	0.121		0.095	0.094	STANDARD OUT OF RANGE
06/26/2011	05:48	FRANKLIN COUNTY SO	JARED J MITCHELL	0.120		0.095		
07/03/2011	02:47	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.083	0.084	0.097	0.097	
07/04/2011	20:44	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.080	0.082	0.097	0.096	
07/14/2011	22:06	FRANKLIN COUNTY SO	JARED J MITCHELL	0.126	0.122	0.096	0.096	
07/15/2011	01:40	FRANKLIN COUNTY SO	CHAD A MILES	0.142		0.096		
07/20/2011	19:21	FRANKLIN COUNTY SO	JEROD J ADAMS	0.148		0.096		
07/24/2011	21:10	FRANKLIN COUNTY SO	COREY L GRISWOLD	R		0.095		
07/28/2011	00:40	FRANKLIN COUNTY SO	COREY L GRISWOLD	0.146	0.144	0.095	0.095	

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/29/2011	02:03	FRANKLIN COUNTY SO	EDOUARD A LARENTE			0.094		STANDARD OUT OF RANGE
07/30/2011	00:32	FRANKLIN COUNTY SO	ERIC W THOMPSON			0.094		STANDARD OUT OF RANGE
08/07/2011	01:09	FRANKLIN COUNTY SO	JARED J MITCHELL			0.089		STANDARD OUT OF RANGE
08/07/2011	01:12	FRANKLIN COUNTY SO	JARED J MITCHELL			0.089		STANDARD OUT OF RANGE
08/27/2011	21:32	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.111		0.095		
10/02/2011	02:36	FRANKLIN COUNTY SO	JEROD J ADAMS	0.186		0.097		
10/15/2011	02:21	FRANKLIN COUNTY SO	JARED J MITCHELL	0.134		0.095		RESPONSE TIMEOUT
10/15/2011	02:27	FRANKLIN COUNTY SO	JARED J MITCHELL	0.130		0.096		
10/16/2011	00:28	FRANKLIN COUNTY SO	ANTHONT M RAINVILLE	0.058	0.056	0.097	0.097	
10/22/2011	22:38	FRANKLIN COUNTY SO	CHAD A MILES	0.264	0.259	0.097	0.097	
11/01/2011	13:41	VDHL	STEVEN E HARNOIS	0.000		0.098		
11/05/2011	23:04	FRANKLIN COUNTY SO	DEAN M SCOTT	0.067		0.097		
11/06/2011	00:38	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			0.096		STANDARD OUT OF RANGE
11/06/2011	00:41	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.096		0.098		
11/11/2011	01:21	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	R		0.097		
11/11/2011	01:42	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.103	0.099	0.097	0.097	
11/12/2011	01:41	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.072		0.097		
11/16/2011	03:03	FRANKLIN COUNTY SO	COREY L GRISWOLD	0.168		0.097		RESPONSE TIMEOUT
12/20/2011	00:07	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			0.095		STANDARD OUT OF RANGE
12/20/2011	00:09	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			0.094		STANDARD OUT OF RANGE
12/22/2011	20:47	FRANKLIN COUNTY SO	JARED MITCHELL			0.096		STANDARD OUT OF RANGE
12/22/2011	20:49	FRANKLIN COUNTY SO	JARED MITCHELL	0.099	0.092	0.097	0.098	
12/29/2011	02:03	FRANKLIN COUNTY SO	JACY L DENNETT	0.146	0.134	0.097	0.097	
01/08/2012	04:05	ST. ALBANS PD	WILLIAM R FURNARI			0.096		STANDARD OUT OF RANGE
01/09/2012	14:28	FRANKLIN COUNTY SO	JAMES R LYNCH	0.000		0.097		
01/10/2012	22:35	FRANKLIN COUNTY SO	JARED J MITCHELL	I		0.098		RESPONSE TIMEOUT
01/27/2012	11:28	FRANKLIN COUNTY SO	JAMES R LYNCH	0.116	0.114	0.098	0.098	
03/16/2012	09:53	NICK HADDEN	DARCY M RICHARDSON	0.000	V	0.099	0.101	
03/16/2012	10:03	NICK HADDEN	DARCY M RICHARDSON	V		0.100		
03/16/2012	10:18	NICK HADDEN	DARCY M RICHARDSON	0.000		0.100		
03/16/2012	10:26	NICK HADDEN	DMR ALCOHOL FREE	0.000		0.100		
03/16/2012	10:43	NICK HADDEN	DARCY M RICHARDSON	0.000		0.100		
03/16/2012	11:08	NICK HADDEN	DARCY M RICHARDSON	0.000		0.100		
03/16/2012	11:13	NICK HADDEN	DARCY M RICHARDSON	0.000		0.101		
03/16/2012	11:34	NICK HADDEN	DARCY M RICHARDSON	0.045	0.046	0.100	0.100	
03/16/2012	11:44	NICK HADDEN	DARCY M RICHARDSON	X	0.047	0.100	0.100	
03/16/2012	11:58	NICK HADDEN	DARCY M RICHARDSON	0.055	0.055	0.100	0.100	
03/16/2012	12:08	NICK HADDEN	DARCY M RICHARDSON	0.044	0.049	0.100	0.101	
03/16/2012	12:38	NICK HADDEN	DARCY M RICHARDSON	0.000		0.100		
03/16/2012	12:44	NICK HADDEN	DARCY M RICHARDSON	0.046	0.044	0.100	0.100	
03/16/2012	12:55	NICK HADDEN	DARCY M RICHARDSON	0.055	0.051	0.100	0.100	
03/16/2012	13:04	NICK HADDEN	DARCY M RICHARDSON	0.051		0.100		
03/16/2012	14:32	NICK HADDEN	DARCY M RICHARDSON	0.008	0.006	0.099	0.100	
03/16/2012	14:40	NICK HADDEN	DARCY M RICHARDSON	0.031	0.026	0.100	0.101	
03/16/2012	15:22	NICK HADDEN	DARCY M RICHARDSON	0.000		0.100		

FCSO St Albans

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
11/25/2010	03:44	Breath Test	RESPONSE TIMEOUT
01/21/2011	03:01	Breath Test	STANDARD OUT OF RANGE
01/21/2011	03:19	Breath Test	STANDARD OUT OF RANGE
02/17/2011	21:46	Breath Test	STANDARD OUT OF RANGE
03/12/2011	04:19	Breath Test	DETECTOR OVERFLOW
03/12/2011	04:26	Breath Test	FILTER 3 WON'T ZERO
03/12/2011	04:29	Breath Test	FILTER 3 WON'T ZERO
03/13/2011	03:07	Breath Test	DETECTOR OVERFLOW
03/13/2011	03:09	Breath Test	DETECTOR OVERFLOW
05/06/2011	20:23	Breath Test	RESPONSE TIMEOUT
05/08/2011	02:25	Breath Test	STANDARD OUT OF RANGE
05/08/2011	02:28	Breath Test	STANDARD OUT OF RANGE
05/28/2011	03:27	Breath Test	STANDARD OUT OF RANGE
05/29/2011	18:04	Breath Test	STANDARD OUT OF RANGE
06/26/2011	05:41	Breath Test	STANDARD OUT OF RANGE
07/29/2011	02:03	Breath Test	STANDARD OUT OF RANGE
07/30/2011	00:32	Breath Test	STANDARD OUT OF RANGE
08/07/2011	01:09	Breath Test	STANDARD OUT OF RANGE
08/07/2011	01:12	Breath Test	STANDARD OUT OF RANGE
10/15/2011	02:21	Breath Test	RESPONSE TIMEOUT
11/06/2011	00:38	Breath Test	STANDARD OUT OF RANGE
11/16/2011	03:03	Breath Test	RESPONSE TIMEOUT
12/20/2011	00:07	Breath Test	STANDARD OUT OF RANGE
12/20/2011	00:09	Breath Test	STANDARD OUT OF RANGE
12/22/2011	20:47	Breath Test	STANDARD OUT OF RANGE
01/08/2012	04:05	Breath Test	STANDARD OUT OF RANGE