DATAMASTER DMT MAINTENANCE LOG

Orange County SD
Department/Agency

102906
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/13/15	**DataMaster Simulator Solution** Lot # 14-10-100 EtOHaq Conc: 0.097 [For RI Range: 0.092 - 0.102 ace Label Here] Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: 2/13/2015	W. Briggs
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Orange County SD
Department/Agency

102906
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/10/14	**DataMaster Simulator Solution** Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: 2/10/14 MPW [For RP ... ace Label Here] RPC	M. Welch / CPT
6/27/14	**DataMaster Simulator Solution** Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 6/27/14 MPW [For RPC ... ce Label Here]	M. Welch / Captain
9/18/14	**DataMaster Simulator Solution** Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 9/18/14 RCD [For RPC ... e Label Here] Solution out of range. Checked simulator and tubing for any leaks. Trimmed tubing at sim. tower fittings to ensure no leaks. Performed RPC and APM	R. Driscoll DMT TECH
12/22/14	**DataMaster Simulator Solution** Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: 12/22/14 MPW [For RPC ... ce Label Here] Solution Change	M Welch / Captain

ALC 803 REV 03 08/2011

DATAMASTER DMT OPERATOR USE LOG

Orange County SD
Department/Agency

102906
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
12/22/14	[illegible]	TEST/[illegible]	[illegible]	
01/16/15	CORLISS	ZAC HAYWARD		
01/16/15	Briggs	Dexter, Ronald	.095	.094
2/13/15	Briggs	RPC / Chuck		

ALC 603 Rev 3 08/11

DATAMASTER DMT OPERATOR USE LOG

Orange County SD
Department/Agency

102906
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
11/5/13	Briggs	Culver, William	.092	.092
11/6/13	Briggs	TEST	.090	
11-8-13	D. Eggum TEST/ Simulator Solution Change	TEST Simulator Solution Change	0.094	
11/10/13	D. BOHNYAK	DANIEL WOOD	0.093	
11/10/13	D. BOHNYAK	DANIEL WOOD	.093	
11/27/13	Clark	Andre Dechamme	out of Range	
11/27/13	TEST D Eggum	TEST OUT of Range		
12/5/13	R. Driscoll	TEST / ALP	.093	—
12/5/13	R. Driscoll	TEST / Solution Change	.097	—
1/5/13	~~KCB~~ D. BOHNYAK	KENNETH HODGE		
1/16/14	Briggs	Merrill, Fayanne	.093	
2/19/14	M. Welch	TEST RPC	.096	—
3/14/14	Briggs	Test	.095	—
5/12/14	Pregent	Timothy Coots	.094	—
6/27/14	Welch, Michael P	TEST/RPC	.097	
8/30/14	BARITEAU JM R	Joshua Perry	.094	
9/15/14	Pierce, Joel	Crawford, Winston	.090	
9/18/14	R. Driscoll	TEST / RPC	.096	—
9/20/14	R. Driscoll	TEST / ARM	.090	—
12/20/14	Bariteau JM R	Grant, Edward	out of Range	

ALC 603 Rev 3 08/11