DATAMASTER DMT OPERATOR USE LOG

Orange County Sheriff
Department/Agency

102106
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4/12/16	Jean-Miguel Barteau	Jonathon Marineau	0.099	wrong Dept Entered
4/12/16	Jean-Miguel Barteau	Jonathon Marineau	Simulator times out	
4/12/16	Jean-Miguel Barteau	Jonathon Marineau	0.096	0.096
4/4/16	Jean-Miguel Barteau	Bryce Washburn	Simulator timed out	
5/4/16	Jean-Miguel Barteau	Bryce Washburn	Simulator time out	
5/4/16	Jean-Miguel Barteau	Bryce Washburn	Simulator tim out	
5/4/16	W. Briggs	Test	.095	
5/9/16	W. Briggs	Accuracy + Precision	Simulator Timed Out.	
5/9/16	W. Briggs	Solution Changed	.097	

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

Orange SD
Department/Agency

102906
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/3/2015	**DataMaster Simulator Solution** Failed Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 10-3-2015 Solution Failed [For R … lace Label Here] **DataMaster Simulator Solution** Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 10-3-2015	Wayne Briggs Deputy
10/6/15	Printer not taking paper, adjusted paper tray and now [For RPC or Simulator Solution Change Place Label Here] working properly. Completed APM protocol. Downloaded data to USB drive, uploaded new software, and installed new I/O PCB. Not currently connected to DPS server.	R. Driscoll ~~TGS~~ KO DMT Tech.
2/3/16	**DataMaster Simulator Solution** Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 2/3/2016 [For … Place Label Here]	Wayne Briggs
5/9/16	Accuracy and Precision Test ~~Solution Failed~~ Simulator Timed Out	Wayne Briggs
5/9/16	**DataMaster Simulator Solution** Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 5/9/2016 [For RPC … e Label Here]	Wayne Briggs

ALC 803 REV 03 08/2011